



Government of Karnataka

BIDAR INSTITUTE OF MEDICAL SCIENCES, BIDAR.

(Autonomous Institution of Govt. of Karnataka)

Affiliated to Rajiv Gandhi University of Health Sciences, Karnataka, Bangalore & Recognized by
Govt. of India, New Delhi, Phone: 08482 228366 / Fax No: 08482 240322.

Website:- www.brims-bidar.in & email:- director@brims-bidar.in

BRIMS/TENDER/ENT/CR- 123/2017-18

Dated: 12.03.2018

Amendment

Subject: E-tender for Supply of ENT Instruments and Equipments for the year 2017-18.

Reference: 1.Tender Notification:-No. BRIMS/TENDER/ENT/CR- 123/2017-18 Dated: 22.02.2018

2. Pre-bid Meeting Held on 8.03.2017

In the Pre-bid meeting held on 08.03.2018 after detail discussion, it has been decided by the tender verification committee of BRIMS to effect certain changes to Technical Specifications prescribed in the tender notification and to issue an amendment accordingly the following changes are made to the Technical Specifications for the tender.

SL	Particulars	Amended as
01	<u>TECHNICAL SPECIFICATION FOR NASAL ENDOSCOPY UNIT:-</u> <u>Page No:43 S N:9</u> <u>Xenon Light Source:</u> Xenon light source of 175 Watts Should be able to produce color temperature of 6000 K Should have continuous manual adjustment of light output. Should have standby mode and automatic recovery of last setting of intensity of light. Should be able to display lamp life in digital form and should give visual indication for replacement of xenon lamp in case of lamp life of 500 Hrs is over. Should be certified IEC 601-1 and CE according to MDD.	Light Source should be LED of 250 watts & Above.
02	<u>Page No:43 S N:10</u> Xenon spare lamp of 175 Watts suitable for Xenon light source 2 No	LED Spare Lamp of 250 watts & Above should be supplied 2 No
03	<u>SPECIFICATIONS FOR ENT MICRODEBRIDER, MASTOID MICROMOTOR AND STAPES DRILL:</u> Page No 45 SN:01 para 14 Should have the provision to mount the console on various sizes of IV pole.	Should have the provision to mount the console on various sizes of IV pole or table top

Director

Bidar Institute of Medical Science

Bidar